

## The determination of oxygen saturation in the corrected nomogram of Siggaard Andersen

De Armelini, M.E.G. and Armelini, Pedro Alberto (1972) *The determination of oxygen saturation in the corrected nomogram of Siggaard Andersen*. Revista Asociación Bioquímica Argentina, 37. pp. 162-166.

El texto completo no está disponible en este repositorio.

### RESUMEN

The values derived from the 'Blood Gas Calculator' for oxygen saturation of the venous blood of adults were significantly higher than the spectrophotometric values. No significant difference was found between the values of oxygen saturation in the capillary blood of adults and neonates. It is therefore possible to use either method for the capillary blood of children or adults.

**TIPO DE DOCUMENTO:** Artículo

**PALABRAS CLAVE:** Blood. Methodology. Nomogram. Oxygen consumption.

**TEMAS:** [R Medicina > R Medicina \(General\)](#)

**UNIDAD ACADÉMICA:** [Universidad Católica de Córdoba > Facultad de Ciencias de la Salud](#)